



FROM EXTERNAL
GALIL RIO ANALOG
OUTPUT

J1
1
6
2
7
3
8
4
9
5
DB-9 , MALE

NOTES:

1. ALL AC WIRING MUST BE 14 AWG OR LARGER.
2. WIRE EACH AC GROUND CONNECTION TO GROUNDING BOLT.
3. MOUNT AC FUSE ON FRONT PANAL.

B

REVISIONS	http://loel.ucolick.org/manual/Apf/schematics/APF_Proportional heater controller1.PDF
04-08-13	CREATED DOCUMENT
10-25-13	INSTALLED HEATER & UPDATED DOCUMENT

UNIVERSITY OF CALIFORNIA LICK OBSERVATORY		PROPORTIONAL AC POWER CONTROLLER	
DES'N BY: M.SAYLOR	ORIGIN DATE: 04-07-13	DWG. NO.	NUM. 1 OF 2
Last SN	MODIFY DATE: 10 25-13	EL-3598	REV. A
PRINT TIME: 16:18:11	PRINT DATE: Fri Oct 25, 2013		
PATH: APF/		Proportional heater controller.sch	