



NOTE: ON EACH AMP-1140, INCOM & LSCOM ARE JUMPERED TO +5V

** (n) = AC SOCKET NUMBER

REVISION
12-17-98 ba COMPLETED DRAWING AND CHANGED REVISION TO B
01-27-99 tc ADDED 3 LAMPS AND A SWITCH. MADE IT A 'B' SIZE SHEET.
02-26-99 tc ADDED QUARTZ LAMP RELAY AND LAMP COOLING FAN
5-27-99 CHANGED QUARTZ LAMP SUPPLY FROM VARIAC TO 15V SUPPLY
6/14/99 CORRECTED PEN-RAY LAMP POWER SUPPLY PART NUMBERS AND REMOVED TEST SWITCH
10-4-99 RELABELED LAMPS AS SET UP DURING COMMISSIONING

UNIVERSITY OF CALIFORNIA LICK OBSERVATORY		CALIBRATION LAMP STAGE WIRING ESI SPECTROGRAPH	
DES'N BY: B Alcott	ORIGIN DATE: 09-12-97	DWG. NO.	NUM. 1 OF 1
Last SN	MODIFY DATE: 10-4-99	EL-3368-2L	
PATH: ES\LMPSTAGE.sch	REV.C		