

## 8-CHANNEL CONTROLLER ANALOG INPUT ASSIGNMENTS

Hamilton Spectrograph Analog Inputs			
CH	SIGNAL	FUNCTION	PIN
1	ANALOG INPUT 1	ON BOARD TEMPERATURE	2
2	ANALOG INPUT 1	(UNASSIGNED)	10
3	ANALOG INPUT 1	(UNASSIGNED)	3
4	ANALOG INPUT 1	24V POWER SUPPLY MONITOR	11
5	ANALOG INPUT 1	-	4
6	ANALOG INPUT 1	-	12
7	ANALOG INPUT 1	-	5
8	ANALOG INPUT 1	-	13

REVISIONS

[http://loel.ucolick.org/manual/Hamilton\\_II/schematics/HAM\\_II\\_in\\_out\\_sh\\_4.pdf](http://loel.ucolick.org/manual/Hamilton_II/schematics/HAM_II_in_out_sh_4.pdf)

UNIVERSITY OF CALIFORNIA LICK OBSERVATORY		ANALOG INPUTS SIGNAL ASSIGNMENTS HAMILTON SPECTROGRAPH II	
DES'N BY: B. Alcott	ORIGIN DATE: 07-11-06	DWG. NO.	NUM. 4 OF 7
Last SN	MODIFY DATE:	EL-4005	REV. A
PRINT TIME: 06:17:15	PRINT DATE: Fri Mar 12, 2010		
PATH: HAMILTON_II\		HAM_II_in_out.sch	